



MICROBIX

REDTMFLOQ[®]

REDxTMFLOQ[®] Influenza A Swab Positive Control



Microbix Biosystems Inc.
265 Watline Avenue
Mississauga, Ontario,
Canada, L4Z 1P3

Cat#: RED-S-13-01

FLOQ[®] is a trademark of Copan Italia Spa



About this package insert

Thank you for your interest in this REDxTM quality control product. This package insert consists of two pages.

- The first page contains the product name and an explanation of the symbols used on the labeling.
- The second page contains the complete package insert text.

If the package insert you view or print does not contain two pages, or if you experience any problems, please email us at customer.service@microbix.com.
By phone: US customers call +1-800-794-6694; International customers call collect +1-905-361-8910.

A printed package insert will be sent to you upon request.

P/N RED-S-13-01.5R0

Explanation of symbols used in Microbix product labeling



Upper limit
of temperature



Temperature
limitation



In Vitro Diagnostic
Medical Device



Single-use only



Positive control



Use By



"Caution, consult
accompanying documents"



Catalogue
number



Batch code



Manufacturer



WARNING: THESE REAGENTS MUST NOT BE SUBSTITUTED FOR THE MANDATORY POSITIVE AND NEGATIVE CONTROL REAGENTS PROVIDED WITH MANUFACTURED TEST KITS.



REDx™ FLOQ® Influenza A Swab Positive Control

FOR IVD USE.

INTENDED USE

REDx™ FLOQ® Influenza A Swab Positive Control is a desiccated, unassayed control intended to monitor laboratory testing performance, procedures, and workflow with molecular and immunodiagnostic assays that detect Influenza A virus in human nasopharyngeal, oropharyngeal, nasal mid-turbinate, anterior nares and lower respiratory samples, collected on swabs.

PRODUCT DESCRIPTION

REDx™ FLOQ® Influenza A Swab Positive Control is formulated with inactivated native Influenza A virus. REDx™ FLOQ® Influenza A Swab Positive Control can be utilized as an external sample to monitor the processes of (1) respiratory virus immunoassays and (2) respiratory virus molecular assays, including sample extraction and purification, amplification, and detection¹.

REDx™ FLOQ® Influenza A Swab Positive Control does not have an assigned value ("unassayed"). Laboratories are required to establish an acceptance range for each lot of REDx™ FLOQ® Influenza A Swab Positive Control with all assay procedures that the control is intended to be used with, prior to routine use in the laboratory^{2,3}.

PRINCIPLES OF THE PROCEDURE

REDx™ FLOQ® Influenza A Swab Positive Control is designed as an external independent sample for use with laboratory testing of Influenza A virus targets, according to ISO15189 and CLIA regulations.

REAGENTS

Cat. No RED-S-13-01; 1 swab containing inactivated Influenza A virus.

LIMITATIONS OF THE PROCEDURE

REDx™ CONTROLS MUST NOT BE SUBSTITUTED FOR THE POSITIVE AND NEGATIVE CONTROL REAGENTS PROVIDED WITH THE MANUFACTURED TEST KITS.

TEST PROCEDURES and INTERPRETATION OF RESULTS provided by manufacturers of test kits must be followed closely.

Deviations from procedures recommended by test kit manufacturers may produce unreliable results. REDx™ FLOQ® Influenza A Swab Positive Control DOES NOT HAVE AN ASSIGNED VALUE and may not be suitable for use with all respiratory viral test kits and procedures. Procedures for implementing a quality assurance program and monitoring test performance on a routine basis must be established by each individual laboratory. Each laboratory should establish its own range of acceptable values^{2,3}.

Controls are not calibrators and should not be used for assay calibration.

Adverse shipping and storage conditions or use of outdated controls may produce erroneous results.

REDx™ FLOQ® Influenza A Swab Positive Control might not be suitable for NATs without an extraction step.

WARNINGS AND PRECAUTIONS

For IVD use.

For Professional and Trained Laboratory Personnel Use Only

Safety Precautions

- Raw material used for REDx™ FLOQ® Influenza A Swab Positive Control preparation is inactivated.
- Use Centers for Disease Control and Prevention (CDC) recommended universal precautions for handling the samples and human specimens⁴.
- REDx™ FLOQ® Influenza A Swab Positive Control must be disposed of by following RCRA ID#D001 guidelines for ignitable waste⁵.
- Keep REDx™ FLOQ® Influenza A Swab Positive Control pouch closed when not in use.

HANDLING PRECAUTIONS

- Do not use controls beyond the expiration date.
- Avoid contamination of controls when opening the swab pouches.

STORAGE INSTRUCTIONS

Store REDx™ FLOQ® Influenza A Swab Positive Control at 2-30°C until use.

Once opened, REDx™ FLOQ® Influenza A Swab Positive Control should not be reused.

MATERIALS PROVIDED

REDx™ FLOQ® Influenza A Swab Positive Control – 1 swab

MATERIALS REQUIRED, BUT NOT PROVIDED

Refer to the instructions supplied by the test kit manufacturer for guidance on how to use the REDx™ FLOQ® Influenza A Swab Positive Control.

PROCEDURE

When including the REDx™ FLOQ® Influenza A Swab Positive Control in a test run, the exact same procedure for unknown specimens collected on a swab must be used. Refer to the manufacturer's supplied instructions for use provided with the respiratory viral test kit.

For Use with Molecular Diagnostic Tests:

- Elute the REDx™ FLOQ® Influenza A Swab Positive Control by referring to the preferred technique and volumes described in the assay's Instructions For Use (usually 1-3 mL).
- For better volume recovery, after incubation, swirl the swab 5-10 times in the vial and remove the swab by pressing it towards the walls of the elution vial.
- Use 100-1000 µL from the eluted REDx™ FLOQ® Influenza A Swab Positive Control for the nucleic acid extraction step.
- After extraction, proceed with the molecular assay by using the eluted nucleic acid test volume specified in the assay's Instructions For Use (usually 5-20 µL from the eluted purified nucleic acid volume).

For Use with Immunodiagnostic Tests:

- Elute the REDx™ FLOQ® Influenza A Swab Positive Control by referring to the preferred technique and volumes described in the assay's Instructions For Use (usually 0.2-0.5 mL).
- For better volume recovery, after incubation, swirl the swab 5-10 times in the vial and remove the swab by pressing it towards the walls of the elution vial.
- Proceed with the detection step by using the test volume specified in the assay's Instruction For Use (usually 50-200 µL from the eluted swab volume).

NOTE: Controls must NOT be substituted for the internal positive and negative controls supplied with the test kit.

As the REDx™ FLOQ® Influenza A Swab Positive Control does not have an assigned value, the laboratory must establish an acceptance range for each lot of REDx™ FLOQ® Influenza A Swab Positive Controls.

TROUBLESHOOTING

When REDx™ FLOQ® Influenza A Swab Positive Control results are outside of the established laboratory acceptance range for internal controls, it may be an indication of unsatisfactory test performance.

Possible sources of error include deterioration of test kit reagents, operator error, faulty performance of equipment, or contamination of reagents; internal laboratory procedures should be followed.

REFERENCES

- Accurate Results in the Clinical Laboratory 2013*, ISBN: 978-0-12-415783-5
- Kinns H, Pitkin S, Housley D, et al. *J Clin Pathol* 2013;66:1027-1032.
- Statistical Quality Sample for Quantitative Measurements: Principles and Definitions; Approved Guideline— Second Edition*. NCCLS document C24-A2, 1999.
- CDC Recommendations for prevention of HIV transmission in health care settings*. *MMWR* 36 (suppl. 2), 1987.
- Treatment standards for hazardous waste; 40 CFR 268.40 Subpart D. D001: Ignitable characteristics of waste*.

For assistance, contact Microbix Technical Support at +1-905-361-8910.



P/N RED-S-13-01.5R0
25 August, 2022

CERTIFICATE OF ANALYSIS

REDxTMFLOQ[®] Influenza A Swab Positive Control

Product Information

Product Name	REDx TM FLOQ [®] Influenza A Swab Positive Control
Catalogue Number	RED-S-13-01
Strain/Cell line	Influenza A/H1N1 Subtype
Lot number	13010019A3
Analyte Value	Unassayed. Value not assigned by the manufacturer
Product Description	Unassayed qualitative positive control for the detection of Influenza A virus.
Elution Volume	1-3 mL
Storage conditions	2-30°C
Expiry Date	07 Dec 2024
Intended Use	For <i>In vitro</i> Diagnostic Use.

Quality Control Information

Test method(s)	Cepheid Xpert [®] Xpress SARS-CoV-2/Flu/RSV Assay Quidel Sofia 2 SARS/Flu Assay BinaxNOW Influenza A+B Card
Test Result(s)	Observed Result
SARS-CoV-2	Negative
Influenza A	Positive
Influenza B	Negative
Respiratory Syncytial Virus (RSV)	Negative
Inactivation	Inactivated

This product is an unassayed control. The reported QC information is not intended to represent product specifications in any commercial or lab developed test. The laboratory should establish its own analyte value and ranges.



Microbix Biosystems Inc.
 265 Watline Avenue
 Mississauga, Ontario, L4Z 1P3



Information

Please ensure that your institution is consulted before recycling an item.



Quality Assurance Signature:

Date: 10 Oct, 2023